



## SOP Title: RT-qPCR for Crimean Congo Haemorrhagic Fever Virus

SOP ID: 315, Version 1.0

Effective Date: 3 March, 2026

Review Date: 3 March, 2027

Primary Author (s): Tom Edwards, Caitlin Greenland-Bews

Secondary Author: Eric Fevre

Approver: Siobhan Mor

### 1. Purpose

To ensure a standardized and consistent approach to the PCR-based detection of **Crimean Congo hemorrhagic fever virus RNA in serum samples** collected as part of the ZAFI project.

### 2. Scope

This SOP applies to all laboratory personnel involved in the processing, analysis, and interpretation of samples for dengue, chikungunya and zika virus PCR testing under the ZAFI project across Kenya, Uganda, and Ethiopia.

### 3. Abbreviations

- **RT-qPCR:** Reverse Transcriptase Quantitative Polymerase Chain Reaction
- **CCHF:** Crimean Congo Haemorrhagic Fever

### 4. Responsibilities

Role	Responsibility
<b>Laboratory technician</b>	Perform laboratory testing according to this SOP; handle and process samples appropriately; adhere to biosafety and quality control procedures; document results accurately in logbooks and KOBO tools; report deviations or issues to the laboratory lead/supervisor.
<b>Laboratory lead/supervisor</b>	Oversee laboratory testing; ensure staff competency and adherence to quality and biosafety standards; review and validate results; escalate issues as appropriate.
<b>Site PI</b>	Ensure relevant staff are trained on this SOP.

### 5. Safety and Precautions

- Preparation of mastermix should be conducted in a “Clean room” or “Clean hood” that is separate to template RNA and any amplification procedures to prevent contamination
- Template (RNA) and samples should be stored separately to mastermix and assay reagents
- When calculating the volume of mastermix to make, include enough volume for a single positive control per target and a negative control

## 6. Consumables and Reagents

Item	
Reagents	Equipment
TaqMan Fast Virus 1-Step Multiplex Master Mix (No Rox)	Microcentrifuge
Primers and Probes	Pipettes (10µl, 20 µl, 200 µl, 1000 µl)
Nuclease Free Water	Vortex
CCHF positive control RNA (gBlock)	<b>Tubes/Plates/other</b>
TaqMan Fast Advanced Master Mix	Plastics consumables
Primers and Probes	Pipette tips
	Disposable gloves

## 7. RT-qPCR Procedure

### 7.1 Preparing the PCR Reaction Mix

- Retrieve the working (10uM) stocks of primers, probes and master mix for the CCHF RT-qPCR and fully thaw reagents on ice.
- When completely thawed vortex each tube briefly and spin in microcentrifuge to collect liquid at the bottom.
- Calculate the number of reactions needed for the samples (one reaction per sample) **including a positive control for each target and a negative control**. Multiply the volume per reaction of each component by the total number of reactions.
- Combine these volumes into one 1.5ml Eppendorf. **\*This should be done in a separate room/hood to template addition.\***

Table of Reactions			
Component	Final concentration	Volume per Reaction	Total volume to add to the 1.5ml Eppendorf assuming $n$ reactions
		96-well fast (0.1ml) plates	
<b>Reaction mix</b>			
TaqMan Fast Virus 1-Step Multiplex Master Mix (No Rox)	1x	2.75	$2.75 \times n$
CCHF Forward Primer (10uM)	900nM	1.24	$1.24 \times n$
CCHF Reverse Primer (10uM)	900nM	1.24	$1.24 \times n$
CCHF Probe (10uM)	625nM	0.86	$0.86 \times n$
NFW	-	4.92	$4.92 \times n$
<b>Aliquot per well</b>			
Reaction mix		10ul	
DNA/Template volume		2.5ul	
<b>Total volume</b>		<b>12.5ul</b>	

- Vortex briefly to mix and spin to ensure all the liquid is at the bottom
- Transfer 10µl of PCR Reaction Mix to each well of the plate
- Retrieve RNA eluates immediately prior to template addition and thaw on ice
- Transfer the plate to template addition room/space and add 2.5µl RNA (extracted from sample), or nuclease-free water for Non-Template Control, to each well. Include Nuclease Free Water as a Non-Template Control and positive RNA from the CCHF gBlock as a positive control.

- i) Seal the reaction plate with optical adhesive film or strip-caps and then centrifuge briefly to bring the PCR reaction mix to the bottom of the well.

### 7.2 Setting up the machine

- a) Set up the following run profile on the qPCR machine.  
 b) Ensure that each well is set to collect data in the FAM channel

Thermal Profile for CCHF qPCR				
Step	Temperature	Time	Cycles	Fluorescence acquisition (channel)
Reverse Transcription	50	5:00	1	
RT-Inactivation/Initial Denaturation	95	0:20	1	
Denature	95	0:15	40	
Anneal/Extend	60	1:00		FAM

### 7.3 Interpretation

- a) The negative control must have no amplification present in any channel  
 b) The positive controls must have amplification in their channel (FAM for CCHF)  
 c) Results with a Ct Below 38 in the FAM channel can be considered positive.

## 8. Record keeping

- Enter results into the results sheet
- Enter results into the dedicated **KOBO form**

## 9. Primer Sequences

Primer/probe ID	Sequence
CCHF Forward	TCTCAAAGAAACACGTGCC
CCHF Reverse	CCTTTTGAACCTTCAAACC
CCHR Probe	(FAM)ACTCAAGGKAACACTGTGGGCGTAAG(BHQ1)

## 10. Review and Approval

Approved By: *Prof Siobhan Mor*

Title: *Chief Investigator, ZAFI*

Date: *3 March 2026*